

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

## ***NivuFlow 750 (incorporating POA-K sensor) Echo Profile Flowmeter***

Manufactured by:

### ***NIVUS GmbH***

Im Täle 2  
75031 Eppingen  
Germany

has been assessed by Sira Certification Service  
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Water Monitoring Equipment,  
Version 2.4 dated February 2013**

Certification Ranges :

Velocity: 0.1m/s to 2.5m/s  
Fluid depth: 0.075m to 0.8m

Project No.: 70046568  
Certificate No: Sira MC160303/00  
Initial Certification: 21 July 2016  
This Certificate issued: 21 July 2016  
Renewal Date: 20 July 2021

Emily Alexander  
Deputy Certification Manager

MCERTS is operated on behalf of the Environment Agency by

## **Sira Certification Service**

Unit 6, Hawarden Industrial Park  
Hawarden, Deeside, CH5 3US  
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.  
For conditions of use, please consider all the information within.  
This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*